

Microcomputer-based artificial vision support system for real-time image processing for camera-driven visual prostheses

Wolfgang Fink

California Institute of Technology
Division of Physics, Mathematics, and Astronomy
Visual and Autonomous Exploration Systems Research
Laboratory
Pasadena, California 91125
and
University of Arizona
Department of Electrical and Computer Engineering
and
Department of Biomedical Engineering
Tucson, Arizona 85721

Cindy X. You

Mark A. Tarbell

California Institute of Technology
Division of Physics, Mathematics, and Astronomy
Visual and Autonomous Exploration Systems Research
Laboratory
Pasadena, California 91125

Abstract. It is difficult to predict exactly what blind subjects with camera-driven visual prostheses (e.g., retinal implants) can perceive. Thus, it is prudent to offer them a wide variety of image processing filters and the capability to engage these filters repeatedly in any user-defined order to enhance their visual perception. To attain true portability, we employ a commercial off-the-shelf battery-powered general purpose Linux microprocessor platform to create the microcomputer-based artificial vision support system (μ AVS²) for real-time image processing. Truly standalone, μ AVS² is smaller than a deck of playing cards, lightweight, fast, and equipped with USB, RS-232 and Ethernet interfaces. Image processing filters on μ AVS² operate in a user-defined linear sequential-loop fashion, resulting in vastly reduced memory and CPU requirements during execution. μ AVS² imports raw video frames from a USB or IP camera, performs image processing, and issues the processed data over an outbound Internet TCP/IP or RS-232 connection to the visual prosthesis system. Hence, μ AVS² affords users of current and future visual prostheses independent mobility and the capability to customize the visual perception generated. Additionally, μ AVS² can easily be reconfigured for other prosthetic systems. Testing of μ AVS² with actual retinal implant carriers is envisioned in the near future. © 2010 Society of Photo-Optical Instrumentation Engineers. [DOI: 10.1117/1.3292012]

Keywords: portable artificial vision system; microprocessor; image processing; visual prosthesis; smartphone; real-time image processing.

Paper 09245R received Jun. 11, 2009; revised manuscript received Oct. 26, 2009; accepted for publication Nov. 24, 2009; published online Jan. 25, 2010.

1 Introduction

Systems providing artificial vision are becoming a reality.¹⁻²³ In particular, (extraocular or intraocular) camera-driven epiretinal implants are being used in chronic patient trials already^{6,5} (Fig. 1). With blind subjects, it is difficult to predict exactly what they can perceive with such camera-driven visual prostheses, especially since they (currently) provide only tens of stimulating retinal electrodes, thereby allowing only for limited visual perception (pixelation). Thus it is important to offer them a wide variety of image processing filters and the capability to engage these filters repeatedly in any user-defined order to enhance their visual perception in daily life.

Image processing systems (Fig. 2), such as the artificial vision simulator (AVS)²⁴⁻²⁶ (Fig. 3), provide these capabilities and perform real-time (i.e., 30 fps) image processing and enhancement of digital camera image streams before they enter the visual prosthesis. AVS, in particular, comprises numerous

efficient image manipulation and processing modules, such as user-defined pixelation, contrast and brightness enhancement, grayscale equalization for luminance control under severe contrast and brightness conditions, user-defined grayscale levels for potential reduction of data volume transmitted to the visual prosthesis, blur algorithms, and edge detection.^{27,28} AVS provides the unique ability and flexibility for visual prosthesis carriers to further customize their individual visual perception afforded by their respective vision systems by actively manipulating parameters of individual image processing filters, even altering the sequence of these filters.

The original implementation of the AVS processing system has been laptop computer based and hence not truly portable. To provide independent mobility for the blind and visually impaired using camera-driven artificial visual prostheses (Fig. 1), we introduce in the following a stand-alone, battery-powered, portable microcomputing platform and software system for real-time image processing for camera-driven artificial vision prostheses, the microcomputer-based artificial vision support system (μ AVS²).²⁹

Address all correspondence to: Wolfgang Fink, PhD, Visual and Autonomous Exploration Systems Research Laboratory, California Institute of Technology, 1200 East California Blvd, Mail Code 103-33, Pasadena, CA 91125. Tel: 626-395-4587; E-mail: wfink@autonomy.caltech.edu




Fig. 1 One instantiation of an artificial vision prosthesis: an intraocular retinal prosthesis using an external microelectronic system to capture and process image data and transmit the information to an implanted microelectronic system. The implanted system decodes the data and stimulates the retina via an electrode array with a pattern of electrical impulses to generate a visual perception.




Fig. 2 Schematic diagram of a real-time image processing system for artificial vision prostheses, which are driven by an external or internal camera system.





Fig. 3 (Top) Typical palette of image processing modules that can be applied in real time to a video camera stream driving an artificial visual prosthesis. (Bottom) a user-defined subset of filters (pixelation plus grayscale equalization plus contrast enhancement) applied in a real world scenario: recognizing a door and doorknob in a low contrast environment.

2 Methods

2.1 Hardware Platform

For the purpose of creating a small, stand-alone, battery-operated, and portable version of AVS, namely μ AVS², we employed a commercial off-the-shelf, battery-powered, general purpose miniaturized Linux processing platform: the Gumstix™ (by Gumstix, Incorporated, Portola Valley, California) microcomputer environment (Fig. 4). It is lightweight (8 g), fast (600-MHz clock speed or better), and equipped with USB, Ethernet, and RS-232 interfaces.

In particular, the Verdex-Pro class Gumstix microcomputer has the following specifications (see <http://www.gumstix.net/Hardware/view/Hardware-Specifications/Verdex-Pro-Specifications/112.html>):

- Marvell® PXA270 CPU with XScale™
- 600-MHz clock speed
- Ethernet
- USB
- RS-232
- dynamic RAM+Flash RAM
- 80×20 mm, 8 g, -25 to 85 °C, 3.5 to 5.0 VDC
- battery powered (includes driving USB cameras and on-board USB/Ethernet/RS-232 interfaces).

2.2 Microcomputer-Based Artificial Vision Support System Processing Architecture

For visual processing purposes, μ AVS² imports raw video frames from a camera connected via its built-in USB port (or




Fig. 4 Gumstix™-based μAVS^2 , a stand-alone, battery-operated, portable, general purpose microcomputing platform for real-time image processing.

alternatively via its built-in Ethernet port, see Fig. 5). μAVS^2 then reduces the resolution of the video frames to match the pixel resolution of the patient's visual prosthesis (i.e., pixelation or downsampling, see Fig. 5). μAVS^2 subsequently processes the downsampled video frames through user-selected image filters in a linear, sequential-loop fashion, resulting in vastly reduced memory and CPU requirements during execution, making the use of a microprocessor possible. The frequency and order of the image processing modules are determined via a predefined, user-selectable "control string" or script without recompilation of the image processing code (Fig. 5). μAVS^2 then issues the processed video frame data over an outbound connection (via RS-232, wired Ethernet, or Internet) to the visual prosthesis system in real time (Fig. 5).

The control string or script defines the order and frequency that the image filters are applied to the input video frames. Each image filter in a specific sequence is applied to the results of the filter before it, in a net cascading effect. Thus, filter 1 is applied to the raw pixelated video frame; filter 2 is then applied to that result, and so on to the last filter, at which time the processed image is ready to be issued for stimulation

of the prosthesis. This cycle repeats anew for each incoming video frame. Before use, the blind subject is tested and the specific sequence of filters that provides him the best results is determined. This sequence is incorporated into the control script, which is then downloaded onto the device. In the current implementation of μAVS^2 , the control script is not changeable by the user; however, implementing μAVS^2 on an advanced embedded platform such as an Apple iPhone could allow the user the selection of a number of prepared control scripts in real time (e.g., one for daytime, one for nighttime, etc.).

In particular μAVS^2 adheres to the following modular processing architecture (Fig. 5).

Stage 1: camera image acquisition. μAVS^2 utilizes access to an "image generator," i.e., a digital video camera. The camera may be local, directly connected via the built-in USB port, or it may be a local or remote IP camera on an accessible network. For a locally connected (hardwired) camera, μAVS^2 will open an exclusive read channel to the device over the USB bus to gain access to the camera's video buffer. For an IP camera, μAVS^2 will access the camera at its IP address, using the access protocol specific to the type of camera in use (e.g., HTTP, UDP).

Stage 2: image capture. Once μAVS^2 has established a connection to a digital video camera, it must extract discrete image frames from the camera. It does this by reading the camera's YUV frame buffer.

Stage 3: custom image processing including pixelation. The standardized image representation is reduced in resolution (pixelated) to match the pixel resolution of the patient's retinal implant electrode array. For example, if the original image representation has a resolution of 640×480 raw pixels, the image is downsampled to the patient's electrode array size, e.g., 16×16 pixels, thus becoming a 256-byte pixel array. This downsampling affords μAVS^2 its real time capability, as it allows for a subsequent dramatic speed-up of otherwise computationally expensive image processing filters that can now be equally efficiently executed on the reduced-size pixel array rather than the full-resolution camera video frames (for more detail see Sec. 2.4). Specialized image processing




Fig. 5 Schematic of μAVS^2 processing architecture.




Fig. 6 Schematic view of Ethernet-based data exchange protocol between μAVS^2 and a power and data telemetry system of an artificial vision prosthesis.³⁰

filters (e.g., contrast, brightness, gray equalization, gray levels, inversion, edge detection, and image blur^{27,28}) are then applied to the resolution-reduced/pixelated image (pixel array) according to a predefined/user-selectable control string or script to enhance its essential characteristics, cleaning up the image before issuance to the retinal implant.

Stage 4: image data transmission. The final stage in μAVS^2 processing causes the processed pixel array to be sent to a power and data telemetry system (Fig. 5) that subsequently outputs the image data, e.g., by means of electrical current pulses on the patient's retinal implant electrode array to generate visual perceptions, i.e., so-called phosphenes.

2.3 Microcomputer-Based Artificial Vision Support System Connectivity

μAVS^2 supports a variety of interfaces that enable its connectivity to input devices such as digital cameras, and to output devices such as artificial vision systems. The interfaces currently supported are as follows.

2.3.1 Universal Serial Bus Camera Interface

μAVS^2 contains a fully supported, general purpose USB port, allowing for the connection of a wide range of USB devices, including many types of USB web cameras, without the need of an external power supply. Alternatively, accessing an IP camera is possible as well, using an HTTP protocol to an IP camera connected to the onboard Ethernet port. When using a YUV-capable USB or IP camera, the camera's grayscale video frame buffer is directly accessible to the μAVS^2 image processing filters.

2.3.2 Ethernet-Based (Transmission Control Protocol/Internet Protocol) Data Exchange Protocol

To integrate the μAVS^2 with artificial vision prostheses, an extensible Ethernet-based (i.e., TCP/IP) data exchange proto-

col (Fig. 6) has been developed (for details see Ref. 26) to facilitate the transfer of the μAVS^2 -manipulated (video) output data to, e.g., power and data telemetry systems for retinal prostheses.³⁰ This Ethernet-based protocol allows for maximum flexibility between systems, and is capable of two-way data transfer as well as status message transmission. The interface protocol is sufficiently flexible, so that it is also possible for a power and data telemetry system to transmit its own data (e.g., measurement data obtained from the vision implant) back to μAVS^2 for feedback control, further processing, and analysis. Additionally, contingency measures are integrated into the protocol, providing for negative acknowledgements and data resend requests, should the need arise in which data is required to be resent.

2.3.3 RS232 Communication Protocol

Using the onboard universal asynchronous receiver/transmitter, it is possible to write the μAVS^2 -manipulated (video) output data to the Gumstix serial port at speeds of up to 921,600 baud, using a RS-232 8N1 format: each data byte is actually nine bits long, i.e., eight data bits and one stop bit.

2.4 Microcomputer-Based Artificial Vision Support System Real-Time Image Processing Capability

The real-time image processing capability of μAVS^2 is enabled by the downsampling of the native resolution of the artificial vision prosthesis camera system before image processing filters are applied. For example, if the original image representation has a resolution of 640×480 raw pixels (i.e., 307,200 pixels), and if the implanted electrode array of an epiretinal vision implant has a dimension of 16×16 stimulating electrodes (i.e., 256 electrodes), then the downsampling/binning of the 640×480 pixel image to 16×16 pixels affords a speed-up factor of 1200 (i.e., 3 orders of magnitude) in subsequent filter processing. This process allows otherwise computationally expensive image processing filters (e.g., blur-

Table 1 Binning/downsampling performance efficiency of μAVS^2 in frames per second for artificial vision prostheses with various electrode array dimensions and a camera frame resolution of 160×120 pixels.

Camera frame resolution	160×120 pixels			
Electrode array dimension	4×4 pixels	8×8 pixels	16×16 pixels	32×32 pixels
Binning performance	370	354	303	265

ring) to be executed in real time (i.e., in excess of 30 fps) without loss of filter validity/efficiency. In the Appendix in Sec. 5 we show mathematically and numerically, for a set of relevant image processing filters, that downsampling first and subsequent filtering of the reduced size image yields the exact or nearly exact same end result (i.e., is commutable or nearly commutable) compared to downsampling after image processing filters have been applied to the full-resolution camera video frames. This speed-up enables the employment of computationally low power, portable, and battery-powered microcomputing platforms to perform real-time image processing within an artificial vision system, thereby affording users enhanced mobility and independence.

3 Results

Table 1 demonstrates the binning/downsampling performance efficiency of μAVS^2 for various artificial vision prostheses with electrode array dimensions ranging from 4×4 to 32×32 . For example, μAVS^2 is capable of binning an in-memory source image (i.e., a frame) of 160×120 pixels down to 8×8 pixels at a rate of 354 frames per second (fps).

Table 2 shows individual and total filter performance efficiencies for various electrode array dimensions, obtained with an in-memory frame of 160×120 pixels. The results show that the application of filters to the workflow impact the processing only marginally. For example, adding a contrast filter to the 8×8 case reduces the frame rate by only 3, from 354 (no filters) to 351 fps. This demonstrates that binning/

downsampling represents the bulk of the processing, whereas the individual filters (or even a sequence of all listed filters) pose a minimal additional processing burden, i.e., the image processing filters are “lightweight.” In practice, the real-time frame rate is much more dependent on the maximum frame rate of the camera used, some of which are limited to 10 fps, such as the camera used for our development.

As far as the actual CPU utilization of the Gumstix is concerned, using a camera frame resolution of 160×120 pixels with a sustained frame rate of 10 fps (including the bidirectional communication between μAVS^2 and the camera to generate the YUV frame buffer) results in only a 10% CPU load (measured using the Unix “top” command). In practice, the CPU load is linearly dependent on the camera frame rate, theoretically allowing for a maximum of 100 fps at a camera frame resolution of 160×120 pixels, whereas many implanted prostheses require only 60 or fewer frames per second for optimal functioning.

Figure 7 displays application of the previous filters to a typical scenario (i.e., a darkened hallway). Here the simulated electrode array dimensions of the artificial vision prosthesis (i.e., binning) are 32×32 .

Furthermore, duration tests with two common battery types were performed. The first test used a single 6-V 2CR5 alkaline battery, resulting in an average duration of 2.7 h. The second test used a NiMH 6-V \times 5-Ah rechargeable battery pack, resulting in an average duration of 12 h.

Table 2 Individual and total filter performance efficiencies (including prior binning/downsampling) of μAVS^2 in frames per second for artificial vision prostheses with various electrode array dimensions and a camera frame resolution of 160×120 pixels.

Binning dimension	4×4 pixels	8×8 pixels	16×16 pixels	32×32 pixels
Contrast filter	369	351	296	244
Brightness filter	368	349	293	233
Gray equalization filter	300	288	253	221
Gray level filter	370	352	301	256
Inversion filter	371	353	302	262
Edge detection filter	368	341	266	170
Image blur filter	367	349	293	236
All filters engaged	302	280	217	137




Fig. 7 Example filters of μAVS^2 applied to a typical scenario (i.e., a darkened hallway). The electrode array dimensions of the artificial vision prosthesis are 32×32 .

The Gumstix natively supports a multiboot versatility. If a microSD memory card is inserted into the Gumstix's card slot, the Gumstix will boot from the system installed on it instead of booting from its own internal flash RAM. Thanks to this capability, we have been able to provide several custom-tailored image filter cascades, each on its own microSD card, such that a change from one setting to another can easily be accomplished by swapping cards. On reboot, μAVS^2 automatically executes the new processing cascade established on the microSD card without further user interaction, and continuously performs real-time image processing.

4 Conclusion

To support existing and future camera-driven artificial vision prostheses, we have devised the microcomputer-based artificial vision support system (μAVS^2): a small, stand-alone, battery-operated, customizable image processing system that utilizes a general purpose miniaturized Linux processing platform (Fig. 4). μAVS^2 provides real-time image processing for artificial vision systems while maintaining portability and thus independence for its users. Moreover, with its general purpose computing capabilities, it can easily be reconfigured to support prosthetic systems beyond artificial vision, such as control of artificial limbs.

Multi-purpose systems, such as smartphones (e.g., Apple iPhone, BlackBerry, or Android phones), would be ideal platforms to host μAVS^2 , as they provide a cell phone link and integrated GPS functionality, in addition to image processing capabilities. It should be noted that at the time of this writing, the application programming interface (API) of the iPhone software development kit (SDK) did not provide for live cam-

era stream access. Such support is anticipated with the next SDK update. Nevertheless, we have proceeded to port a static version of μAVS^2 to the iPhone, applying the suite of filters to still-image manipulation and storage functionalities. This will be updated to perform real-time video processing on the iPhone as soon as its SDK supports live camera stream access.

We would like to emphasize that the employment of μAVS^2 in visual prosthetics is by no means limited to retinal implants (epi- or subretinal) only.^{1,4-6} On the contrary, μAVS^2 is directly and immediately applicable to any (artificial) vision-providing/stimulating system that is based on an external (e.g., eyeglass-mounted) or internal (e.g., intraocular camera) video-camera system as the first step in the stimulation/processing cascade, such as optic nerve implants,¹⁰⁻¹² cortical implants,¹³⁻¹⁶ electric tongue stimulators,¹⁷⁻²⁰ and tactile stimulators (both electrical and mechanical²¹⁻²³). In addition, μAVS^2 can interface to infrared (IR) camera systems to augment the visual cues with thermal information, allowing for "supervision" at night and during adverse weather conditions such as fog.

Acknowledgments

The work described in this publication was carried out at the California Institute of Technology under support of the National Science Foundation grant EEC-0310723. Fink and Tarbell may have proprietary interest in the technology presented here as a provisional patent has been filed on behalf of Caltech. You has no proprietary interest.

Appendix

In the following we show mathematically and numerically, for a set of relevant image processing filters, that downsampling first and subsequent filtering of the reduced size image yields the exact or nearly exact same end result compared to downsampling after image processing filters have been applied to the full-resolution camera video frames. The degree of commutation between both procedures for the respective image processing filters is reported in the following in units of gray value differences.

1 General Definitions for Image Processing Filter Proofs

In the following, all divisions in $[a/b]$ form denote integer divisions, and all divisions not enclosed within brackets represent regular divisions. All q (quotients, with $0 \leq q \leq 255$ as applied to filters operating on 256 grayscale values) and r (respective remainders) are $\in \mathbb{N}$. Note that in general, for any integer division $[a/b]=q_*$ with corresponding remainder r_* , we can eliminate the integer division notation via

$$q_* + \frac{r_*}{b} = \frac{a}{b}, \quad 0 \leq r_* < b.$$

Let x_i , $0 \leq x_i \leq 255$, denote the gray value of the pixel to be filtered, $f(x_i)$ be the image processing filter applied to the pixel, and n denote the number of pixels to be processed into one superpixel, i.e., downsampling. The downsampling is as follows:

$$\bar{x} = \left\lfloor \frac{\sum_{i=1}^n x_i}{n} \right\rfloor = \frac{\left(\sum_{i=1}^n x_i \right) - r_1}{n},$$

with r_1 being the remainder of $\sum_{i=1}^n x_i/n$, $0 \leq r_1 < n$, and q_1 the quotient, $0 \leq q_1 \leq 255$.

2 General Observations

Let $a, b, c \neq 0$, and $\in \mathbb{Z}$, but b/c is not an integer. Then $\lfloor a + b/c \rfloor$ can be rewritten by:

$$\text{If } a > 0 \quad b/c > 0, \text{ then } a + \left\lfloor \frac{b}{c} \right\rfloor$$

$$a < 0 \quad b/c < 0,$$

$$a > 0 \quad b/c < 0, \quad a + \left\lfloor \frac{b}{c} \right\rfloor - 1$$

$$a < 0 \quad b/c > 0, \quad a + \left\lfloor \frac{b}{c} \right\rfloor + 1.$$

With the previous definitions, the downsampling first versus filtering first procedures are:

Procedure1: downsample first, filter second.

$$f\left(\left\lfloor \frac{\sum_{i=1}^n x_i}{n} \right\rfloor\right) = f(\bar{x}).$$

Procedure2: filter first, downsample second.

$$\left\lfloor \frac{\sum_{i=1}^n f(x_i)}{n} \right\rfloor = \overline{f(x)}.$$

2.1 Type 1: Additive Filters

Filter description

$$f(x_i) = x_i + a, \quad a \text{ is a constant } \in \mathbb{Z}.$$

Evaluation

Procedure1:

$$f(\bar{x}) = \bar{x} + a.$$

Procedure2:

since $r_1/n < 1$, and a, q_1 are both integers,

$$\overline{f(x)} = \left\lfloor \frac{\sum_{i=1}^n (x_i + a)}{n} \right\rfloor,$$

$$= \left\lfloor \bar{x} + \frac{r_1}{n} + a \right\rfloor,$$

$$= \bar{x} + a.$$

Comparison

$$f(\bar{x}) = \bar{x} + a = \overline{f(x)}.$$

The order of downsampling and filtering is inconsequential with respect to the resulting gray values. Both procedures are commutable.

2.2 Type 2: Multiplicative Filters

Filter description

$$f(x_i) = \lfloor ax_i \rfloor, \quad a = \frac{b}{d} \text{ is a rational constant } \in \mathbb{R} > 0.$$

Evaluation

Procedure1: let r_2 be the remainder of $f(\bar{x}) = \lfloor a\bar{x} \rfloor$, such that

$$f(\bar{x}) = \left\lfloor \frac{b \left(\left(\sum_{i=1}^n x_i \right) - r_1 \right)}{dn} \right\rfloor,$$

$$= \frac{b \left(\left(\sum_{i=1}^n x_i \right) - r_1 \right) - r_2}{dn},$$

$$= \frac{b \left(\sum_{i=1}^n x_i \right)}{dn} - \frac{r_1}{n} - \frac{r_2}{dn}.$$

Procedure2: let r_i be the remainder of $\lfloor bx_i/d \rfloor$, $0 \leq r_i < d$, such that

$$\overline{f(x)} = \left\lfloor \frac{\sum_{i=1}^n \lfloor ax_i \rfloor}{n} \right\rfloor = \left\lfloor \frac{\sum_{i=1}^n \left(\frac{bx_i - r_i}{d} \right)}{n} \right\rfloor.$$

with $r_3, 0 \leq r_3 < n$, being the remainder of the prior division, it follows

$$\overline{f(x)} = \frac{\left[\sum_{i=1}^n \left(\frac{bx_i - r_i}{d} \right) \right] - r_3}{n},$$

$$= \frac{b \left(\sum_{i=1}^n x_i \right)}{dn} - \frac{\sum_{i=1}^n r_i}{dn} - \frac{r_3}{n}.$$

Comparison

The two procedures differ by

$$\left(\frac{\sum_{i=1}^n r_i}{dn} + \frac{r_3}{n} \right) - \left(\frac{r_1}{n} + \frac{r_2}{n} \right).$$

Note that $0 \leq r_1 < n$, $0 \leq r_i < d$, and $0 \leq r_3 < n$. Hence, the greatest difference with respect to the resulting gray values between the two procedures is

$$\left(\frac{\sum_{i=1}^n r_i}{dn} + \frac{r_3}{n} \right) - \left(\frac{r_1}{n} + \frac{r_2}{n} \right) < 2.$$

Both procedures are nearly commutable.

2.3 Type 3: Inversion Filter

Filter description

$$f(x_i) = 255 - x_i.$$

Evaluation

Procedure1:

$$f(\bar{x}) = 255 - \bar{x}.$$

Procedure2:

$$\overline{f(x)} = \left\lfloor \frac{\sum_{i=1}^n (255 - x_i)}{n} \right\rfloor,$$

$$= \left\lfloor 255 - \frac{\sum_{i=1}^n x_i}{n} \right\rfloor,$$

$$= 255 - \left\lfloor \frac{\sum_{i=1}^n x_i}{n} \right\rfloor - 1,$$

$$= 255 - \bar{x} - 1.$$

Comparison

$$f(\bar{x}) = 255 - \bar{x} = \overline{f(x)} + 1.$$

Therefore, the respective gray values of the two procedures differ by a constant value of 1. Hence, they are nearly commutable.

2.4 Type 4: Gray-Value Reduction Filter

Filter description

Let v denote the new number of gray levels. For practical purposes, $1 < v < 256$, $v \in \mathbb{N}$. Let c be the number of original

grayscale values incorporated into each new grayscale value, and Δ be the interval between gray values of the new grayscale.

$$c = \left\lfloor \frac{256}{v} \right\rfloor, \quad \Delta = \left\lfloor \frac{255}{v-1} \right\rfloor, \quad 1 \leq c \leq \Delta \leq 2c (*).$$

The gray-value reduction filter function then reads:

$$f(x_i) = \left\lfloor \frac{x_i}{c} \right\rfloor \Delta.$$

Evaluation

Procedure1: let r_4 be the remainder of \bar{x}/c , $0 \leq r_4 < c$. Recall that $\bar{x} = ((\sum_{i=1}^n x_i) - r_1)/n$. Hence, Procedure1 yields:

$$f(\bar{x}) = \left\lfloor \frac{\bar{x}}{c} \right\rfloor \Delta = \left(\frac{\bar{x} - r_4}{c} \right) \Delta,$$

$$= \left\lfloor \frac{\left(\frac{\sum_{i=1}^n x_i}{n} - r_1 \right) - r_4}{c} \right\rfloor \Delta,$$

$$= \frac{\Delta \sum_{i=1}^n x_i}{cn} - \frac{\Delta r_1}{cn} - \frac{\Delta r_4}{c}.$$

Procedure 2: let r_i be the remainder of $\lfloor x_i/c \rfloor$, $0 \leq r_i < c$, and r_5 be the remainder of $\lfloor \sum_{i=1}^n \lfloor x_i/c \rfloor \Delta / n \rfloor$, $0 \leq r_5 < v$.

$$\overline{f(x)} = \left\lfloor \frac{\sum_{i=1}^n \left\lfloor \frac{x_i}{c} \right\rfloor \Delta}{n} \right\rfloor,$$

$$= \frac{\sum_{i=1}^n \left\lfloor \frac{x_i}{c} \right\rfloor \Delta - r_5}{n},$$

$$= \frac{\sum_{i=1}^n \left(\frac{x_i - r_i}{c} \right) \Delta - r_5}{n},$$

$$= \frac{\Delta \sum_{i=1}^n x_i}{cn} - \frac{\Delta \sum_{i=1}^n r_i}{cn} - \frac{r_5}{n}.$$

Comparison

The two procedures differ by

$$\left(\frac{\Delta r_1}{cn} + \frac{\Delta r_4}{c}\right) - \left(\frac{\Delta \sum_{i=1}^n r_i}{cn} + \frac{r_5}{n}\right).$$

Recall that $0 \leq r_1 < n$, $0 \leq r_4 < c$, $0 \leq r_i < c$, $0 \leq r_5 < n$, $1 \leq c \leq \Delta \leq 2c$.

$$0 \leq \left(\frac{\Delta r_1}{cn} + \frac{\Delta r_4}{c}\right) < \left(\frac{\Delta}{c} + \Delta\right) < 2 + 2c,$$

$$0 \leq \left(\frac{\Delta \sum_{i=1}^n r_i}{cn} + \frac{r_5}{n}\right) < \left(\frac{\Delta \sum_{i=1}^n c}{cn} + 1\right) = \Delta + 1 < 2c + 1.$$

Therefore, since

$$0 \leq \left(\frac{\Delta r_1}{cn} + \frac{r_4}{c}\right) < 2 + 2c, \quad 0 \leq \left(\frac{\sum_{i=1}^n r_i}{cn} + \frac{r_5}{n}\right) < 2c + 1,$$

$$\left(\frac{\Delta r_1}{cn} + \frac{r_4}{c}\right) - \left(\frac{\sum_{i=1}^n r_i}{cn} + \frac{r_5}{n}\right) < 2c + 2.$$

Thus, the greatest difference between the resulting gray values of the two procedures is $2c+2$. Hence, the lower the number of new gray values, the larger the deviation of resulting gray values, and vice versa. Thus, the degree of commutation between both procedures depends on the number of new gray values.

(*) *Proof of $1 \leq c \leq \Delta \leq 2c$*

Consider $c = \lfloor 256/v \rfloor \geq 1$, $\Delta = \lfloor 255/v - 1 \rfloor = \lfloor (255v/v - 1)/v \rfloor$. With $1 < v < 256$, $v \in \mathbb{N}$, the numerator of Δ can take on values

$$\frac{255^2}{254} \leq \left(\frac{255v}{v-1}\right) \leq \frac{255 * 2}{1}.$$

Note the left side of this inequality is >256 but <257 and corresponds to $v=255$, while the right side corresponds to $v=2$, and is equal to 510. To the left side of this series of inequalities, we can then tag

$$256 < \frac{255^2}{254} \leq \left(\frac{255v}{v-1}\right) \leq \frac{255 * 2}{1},$$

which then implies that $c < \Delta$.

Furthermore, $2c = 2 \lfloor 256/v \rfloor \geq \lfloor 2 * 256/v \rfloor - 1 = \lfloor 512 - v/v \rfloor$. The smallest possible value for the numerator of $2c$ is $512 - v = 512 - 2 = 510 = 255 * 2/1$. Hence, $\Delta \leq 2c$.

2.5 Type 5: Blur Filter

For the blurring, we developed and implemented a computationally inexpensive code that consists of two subsequent sweeps across the respective image (downsampled or full resolution): first a vertical sweep (i.e., by columns) followed by

a horizontal sweep (i.e., by rows). In both sweeps only $NN=1$ horizontal or vertical nearest neighbors, respectively, are considered. The nearest neighbors are equally weighted. This blur procedure delivers an effective blur (e.g., Fig. 7) and is applied in lieu of computationally expensive Gaussian blur filters.^{27,28} We randomly generated 320×320 pixel images, then proceeded to: 1. blur the image once with the number of nearest neighbors $NN=1$, and downsample the image to either 4×4 , 8×8 , 16×16 , or 32×32 ; or 2. downsample the image to either 4×4 , 8×8 , 16×16 , or 32×32 , then blur it. Results between the same downsampling dimensions of the two procedures were compared, and the respective average deviation per pixel was recorded. The numerical simulations revealed that the deviation between the two procedures is minimal (i.e., they are nearly commutable): worst average deviation was 4.4 ± 3.4 on 32×32 downsampled image, and best average deviation was 0.58 ± 0.58 on a 4×4 downsampled image. 10,000 simulation runs were conducted, respectively.

2.6 Type 6: Gray Equalization Filter

For the gray equalization we employed the filter described in Refs. 27 and 28. We randomly generated 320×320 pixel images, then proceeded to: 1. equalize the image once with slope=0, and downsample the image to either 4×4 , 8×8 , 16×16 , or 32×32 ; or 2. downsample the image to either 4×4 , 8×8 , 16×16 , or 32×32 , then equalize it (with slope = 0, see Refs. 27 and 28). Results between the same downsampling dimensions of the two procedures were compared, and the respective average deviation per pixel was recorded. Average deviations were obtained via averaging the differences between corresponding pixels in images produced through first equalization then binning and first binning then equalization. Increased binning results in increased deviation: 32×32 resulted in an average deviation of 58 ± 33 , 16×16 resulted in 61 ± 35 , 8×8 resulted in 64 ± 37 , and 4×4 resulted in 72 ± 34 . 10,000 simulation runs were conducted, respectively. The numerical simulations revealed that large downsampling (e.g., 4×4) after equalization merely centered pixel gray values around the middle of the grayscale (i.e., 128), suggesting that it is more effective to perform gray equalization after downsampling to attain meaningful equalization effects, as it utilizes the entire grayscale range of 0 to 255.

References

1. E. Zrenner, "Will retinal implants restore vision?," *Science* **295**(5557), 1022-1025 (2002).
2. S. C. DeMarco, "The architecture, design, and electromagnetic and thermal modeling of a retinal prosthesis to benefit the visually impaired," PhD Thesis, North Carolina State Univ. (2001).
3. E. Zrenner, K. D. Miliczek, V. P. Gabel, H. G. Graf, E. Guenther, H. Haemmerle, B. Hoefflinger, K. Kohler, W. Nisch, M. Schubert, A. Stett, and S. Weiss, "The development of subretinal microphotodiodes for replacement of degenerated photoreceptors," *Ophthalmic Res.* **29**, 269-28 (1997).
4. J. F. Rizzo, J. L. Wyatt, "Prospects for a visual prosthesis," *Neuroscientist* **3**(4), 251-262 (1997).
5. M. S. Humayun, J. Weiland, G. Fujii, R. J. Greenberg, R. Williamson, J. Little, B. Mech, V. Cimmarusti, G. van Boemel, G. Dagnelie, and E. de Juan Jr., "Visual perception in a blind subject with a chronic microelectronic retinal prosthesis," *Vision Res.* **43**(24), 2573-2581 (2003).

6. W. Liu and M. S. Humayun, "Retinal prosthesis," *IEEE Intl. Solid-State Circuits Conf. Digest Tech. Papers*, pp. 218–219 (2004).
7. P. R. Singh, W. Liu, M. Sivaprakasam, M. S. Humayun, and J. D. Weiland, "A matched biphasic microstimulator for an implantable retinal prosthetic device," *Proc. IEEE Intl. Symp. Circuits Syst.* **4**, 1–4 (2004).
8. J. D. Weiland, W. Fink, M. Humayun, W. Liu, D. C. Rodger, Y. C. Tai, and M. Tarbell, "Progress towards a high-resolution retinal prosthesis," *Conf. Proc. IEEE Eng. Med. Biol. Soc.* **7**, 7373–7375 (2005).
9. J. D. Weiland, W. Fink, M. S. Humayun, W. Liu, W. Li, M. Sivaprakasam, Y. C. Tai, and M. A. Tarbell, "System design of a high resolution retinal prosthesis," *Conf. Proc. IEEE IEDM 2008*, doi: (2008).
10. C. Veraart, C. Raftopoulos, J. T. Mortimer, J. Delbeke, D. Pins, G. Michaux, A. Vanlierde, S. Parrini, and M. C. Wanet-Defalque, "Visual sensations produced by optic nerve stimulation using an implanted self-sizing spiral cuff electrode," *Brain Res.* **813**, 181–186 (1998).
11. C. Veraart, J. Delbeke, M. C. Wanet-Defalque, A. Vanlierde, G. Michaux, S. Parrini, O. Glineur, M. Verleysen, C. Trullemans, and J. T. Mortimer, *IFESS'1999 Proc. Intl. Functional Electric. Stim. Society*, pp. 57–59 (1999).
12. C. Veraart, M. C. Wanet-Defalque, B. Gerard, A. Vanlierde, and J. Delbeke, "Pattern recognition with the optic nerve visual prosthesis," *Artif. Organs* **27**, 996–1004 (2003).
13. W. H. Dobelle, M. G. Mladejovsky, and J. P. Girvin, "Artificial vision for the blind: electrical stimulation of visual cortex offers hope for a functional prosthesis," *Science* **183**(4123), 440–444 (1974).
14. W. H. Dobelle, M. G. Mladejovsky, and J. P. Girvin, "Artificial vision for the blind by electrical stimulation of the visual cortex," *Neurosurgery* **5**(4), 521–527 (1979).
15. W. H. Dobelle, "Artificial vision for the blind: the summit may be closer than you think," *ASAIO J.* **40**(4), 919–922 (1994).
16. W. H. Dobelle, "Artificial vision for the blind by connecting a television camera to the visual cortex," *ASAIO J.* **46**(1), 3–9 (2000).
17. P. Bach-y-Rita, K. A. Kaczmarek, M. E. Tyler, and M. Garcia-Lara, "Form perception with a 49-point electrotactile stimulus array on the tongue: a technical note," *J. Rehabil. Res. Dev.* **35**(4), 427–430 (1998).
18. K. A. Kaczmarek, P. Bach-y-Rita, and M. E. Tyler, "Electrotactile pattern perception on the tongue," *BMES*, pp. 5–131 (1998).
19. R. Kupers and M. Ptito, "Seeing through the tongue: cross-modal plasticity in the congenitally blind," *Intl. Congress Series* **1270**, 79–84 (2004).
20. M. Ptito, S. Moesgaard, A. Gjedde, and R. Kupers, "Cross-modal plasticity revealed by electrotactile stimulation of the tongue in the congenitally blind," *Brain* **128**(3), 606–614 (2005).
21. C. C. Collins and F. A. Saunders, "Pictorial display by direct electrical stimulation of the skin," *J. Biomed. Sys.* **1**, 3–16 (1970).
22. C. C. Collins, "On mobility aids for the blind," in *Electronic Spatial Sensing for the Blind*, D. H. Warren and E. R. Strelow, Eds., pp. 35–64, Martinus Nijhoff, Dordrecht, The Netherlands (1985).
23. K. Kaczmarek, P. Bach-y-Rita, W. J. Tompkins, and J. G. Webster, "A tactile vision-substitution system for the blind: computer-controlled partial image sequencing," *IEEE Trans. Biomed. Eng. BME-32*(8), 602–608 (1985).
24. W. Fink and M. Tarbell, "Artificial vision simulator (AVS) for enhancing and optimizing visual perception of retinal implant carriers," *Invest. Ophthalmol. Visual Sci.* **46**, E-Abstract 1145 (2005).
25. W. Liu, W. Fink, M. Tarbell, and M. Sivaprakasam, "Image processing and interface for retinal visual prostheses," *ISCAS 2005 Conf. Proc.* **3**, 2927–2930, DOI (2005).
26. W. Fink, M. A. Tarbell, L. Hoang, and W. Liu, "Artificial vision support system (AVS²) for enhancing and optimizing low-resolution visual perception for visual prostheses," (to be submitted).
27. J. C. Russ, *The Image Processing Handbook*, CRC Press, Boca Raton, FL (2002).
28. H. R. Myler and A. R. Weeks, *The Pocket Handbook of Image Processing Algorithms in C*, Prentice Hall PTR, Englewood Cliffs, NJ (1993).
29. W. Fink and M. Tarbell, "μAVS²: microcomputer-based artificial vision support system for real-time image processing for camera-driven visual prostheses," *Invest. Ophthalmol. Visual Sci.* **50**, E-Abstract 4748 (2009).
30. G. Wang, W. Liu, M. Sivaprakasam, and G. A. Kendir, "Design and analysis of an adaptive transcutaneous power telemetry for biomedical implants," *IEEE Trans. Circuits Syst.* **52**, 2109–2117 (2005).